

# Magnetically Induced, Visual Detection of Trace Arsenic Contaminants in Water Using Fe<sub>3</sub>O<sub>4</sub> Photonic Crystal Structures

Jain, Raina (School: Greenwich High School)

Arsenic, a highly toxic metal contaminant commonly found in our drinking water, is responsible for many accidental deaths. Currently, the only visual arsenic-in-water detection system is tedious, and can detect arsenic concentrations of 250ppb or more, well above the EPA 10ppb water-action-level. To combat arsenic drinking water contamination, a sensitive, inexpensive, portable, and easily-visualized detection system is needed and has been developed in this research. To begin, superparamagnetic, SiO<sub>2</sub>-coated, polyacrylic acid-capped Fe<sub>3</sub>O<sub>4</sub> colloidal nanocrystals (CNC's) were synthesized, and their photonic and physical properties characterized via SEM and UV-Visible spectroscopy. Application of 80-140G magnetic field from a portable, 3" magnet altered the refractive indices of the photonic structures, so that long-to-short wavelength, red-to-blue color change is easily visualized from the native brown CNC solution color. Addition of 1ml of 10ppb Arsenic, however, to 2ml of 8mg/ml CNCs causes alteration of the photonic characteristics so that long-wavelength shift occurs with an applied magnetic field (native brown to orange). This new color changing behavior is specific to the arsenic contaminant, and attributed to As-O interactions at the surface of the SiO<sub>2</sub>-coated CNCs. Other typical metal contaminants did not share this same metal-oxide CNC-coating affinity. For the consumer-friendly, rapid Arsenic-in-water assay, a color code was developed to detect/indicate as little as 10ppb As-contaminant, with a color change at 10ppb increments. In the field, drops of suspect water are added to the CNC solution at 1:2 (v/v) in a small vial; Arsenic contamination is determined in seconds via color change through the application of a small magnet.

## Awards Won:

Intel ISEF Best of Category Award of \$5,000

First Award of \$5,000

Spectroscopy Society of Pittsburgh: Second Award of \$1000